



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

CAMERON INSTRUMENTS INC.
 173 Woolwich St. Unit 103
 Guelph, ON N1H 3V4
 CANADA
 Julie Grayson Phone: 519 824 7111

CALIBRATION

Valid To: January 31, 2019

Certificate Number: 4171.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ^{2,3,4} (±)	Comments
Gauge Pressure – Gas	(5 to 350) kPa (25 to 1750) kPa (100 to 7000) kPa	0.0025 % + 0.13 Pa 0.0024 % + 0.67 Pa 0.0027 % + 2.6 Pa	Fluke PG7601
	(2000 to 70 000) kPa	0.0042 % + 1.9 Pa	Fluke PG7202
Oil	(2000 to 70 000) kPa (5000 to 200 000) kPa	0.0042 % + 35 Pa 0.0044 % + 19 Pa	Fluke PG7202
Absolute Pressure – Gas	(5 to 350) kPa (25 to 1750) kPa (100 to 7000) kPa	0.0024 % + 0.32 Pa 0.0024 % + 0.72 Pa 0.0027 % + 2.6 Pa	Fluke PG7601
	(2000 to 70 000) kPa + atmospheric pressure	0.0042 % + 2.3 Pa	Fluke PG7202 and Fluke RPM4

Parameter/Equipment	Range	CMC ^{2, 3, 4} (\pm)	Comments
Absolute Pressure – (cont)			
Oil	(2000 to 70 000) kPa + atmospheric pressure (5000 to 200 000) kPa + atmospheric pressure	0.0042 % + 35 Pa 0.0044 % + 19 Pa	Fluke PG7202 and Fluke RPM4

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, percentages are to be read as percent of reading, unless otherwise noted.

⁴ The contributions from the "best existing device", such as resolution and repeatability of the unit under calibration, are not included in the CMC uncertainties claim."





Accredited Laboratory

A2LA has accredited

CAMERON INSTRUMENTS INC.

Guelph, Ontario, CANADA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).



Presented this 14th day of December 2016.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 4171.01
Valid to January 31, 2019

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.